






**Table 2.** X-ray angle values.

		Average	Min	Max	STD. DEV.	Mean std error	Difference	STD.DEV.	Mean std error	Significance (p<0.05)
CPT	Pre	16.06	8.00	27.00	3.00	0.270	3.77	2.38	0.214	0.000
	End	19.83	11.00	33.00	4.15	0.373				
LTC	Pre	38.40	30.00	54.00	4.88	0.403	-7.49	4.03	0.362	0.000
	End	30.91	22.00	41.00	3.44	0.309				
MA	Pre	8.40	1.00	21.00	3.73	0.335	-5.05	3.25	0.292	0.000
	End	3.35	0.00	18.00	3.02	0.271				
KA	Pre	25.14	13.00	39.00	5.43	0.487	-4.86	3.61	0.324	0.000
	End	20.28	13.00	33.00	4.43	0.398				
T-M1	Pre	8.35	1.00	21.00	3.46	0.311	-3.34	3.24	0.291	0.000
	End	5.01	0.00	13.00	2.69	0.242				

**Figure 2.** Comparison of clinical score averages in the pre- and postoperative periods.

Further long-term or prospective randomized studies, preferably comparing different devices, are advisable.

## References

- Bernasconi A, Lintz F, Sadile F. The role of arthroereisis of the subtalar joint for flatfoot in children and adults. *Foot Ankle* 2017;2:438-46.
- Lamm BM, Stasko PA, Gesheff MG, Bhave A. Normal foot and ankle radiographic angles, measurements, and reference points. *J Foot Ankle Surg* 2016;55:991-8.
- Flores D, Mejia Gomez C, Fernandez Hernando M, et al. Adult acquired flatfoot deformity: anatomy, biomechanics, staging, and imaging findings. *Radiographics* 2019;39:1437-60.
- Aronson J, Nunley J, Frankovitch K. Lateral talocalcaneal angle in assessment of subtalar valgus: Follow-up of seventy grice-green arthrodeses. *Foot Ankle* 1983;4:56-63.
- Lau BC, Allahabadi S, Palanca A, Oji D. Understanding radiographs measurements used in foot and ankle surgery. *J Am Acad Orthopaedic Surg* 2022;30:139-54.
- Memeo A, Verdoni F, Rossi L, et al. Flexible juvenile flat foot surgical correction: a comparison between two techniques after ten years' experience. *J Foot Ankle Surg* 2019;58:203-7.
- Indino C, Villafañe JH, D'Ambrosi R, et al. Effectiveness of subtalar arthroereisis with endorthesis for pediatric flexible up at skeletal maturity. *Foot Ankle Surg* 2020;26:98-104.

## Conclusions

The short-time results of Pit'Stop® device in the treatment of juvenile flexible flatfoot are promising: PEEK device should be considered as a valid option.